



WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0035447-1
Request Date 9/25/2023 7:04:00 AM
Schedule Date 9/25/2023 2:00:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1019
 SUP1019
 SUPERIOR MATERIALS 4913 (14)
 1384 LAKE LANSING RD
 LANSING, MI,
 (248) 788-8000

Customer P.O 6300425989
Requester DAN THOMAS
Phone (248) 788-8000
Email dan.hargrave-thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK ONSITE AT 2pm	
WORK PERFORMED	
CALL COMPLETE	
Inspected both scales, performed scale check. Everything looks good.	

Service Date 9/25/2023

2 hours minimum portal to portal

TECHNICIAN(S)
LUCAS SULLIVAN, MATT PLEVA

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs x 20 lbs	341agg	341agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 lbs x 10 lbs	341cem	341cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

Name:

WORK ORDER

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0035447-1
Request Date 9/25/2023 7:04:00 AM
Schedule Date 9/25/2023 2:00:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1019
 SUP1019
 SUPERIOR MATERIALS 4913 (14)
 1384 LAKE LANSING RD
 LANSING, MI,
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8000
Email dan.hargrave-
 thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK ONSITE AT 2pm	

Service Date _____

2 hours minimum portal to portal

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE Inspected both scales, performed scale check. Everything looks good.	LUCAS SULLIVAN, MATT PLEVA

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs x 20 lbs	341agg	341agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 lbs x 10 lbs	341cem	341cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature



Dave



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4913 (14)
 SUP1019 1384 LAKE LANSING RD

Certificate ID: 341cem230927075457

Test Date: 09/25/2023

LANSING, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 341cem Capacity: 10000 lb Graduation: 10 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
09/25/2023	03/26/2024	Semi-Annually	341cem	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM lb	INCREMENT 10

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
5000	lb	5000	lb	0	lb	5000	lb	5000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied		Reading		Error	
3500	lb	5000	lb	8500	lb	0
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4913 (14)
 SUP1019 1384 LAKE LANSING RD
 LANSING, MI

Certificate ID: 341cem230927075457
Test Date: 09/25/2023
ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	17
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	65
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

N/A

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1502	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1503	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1504	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1505	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				

Technician(s): LUCAS SULLIVAN;MATT PLEVA;

Signature: Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 4913 (14)
 SUP1019 1384 LAKE LANSING RD
 LANSING, MI

Certificate ID: 341agg230927075239
Test Date: 09/25/2023
ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	40
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	Description
GL1501	MI--03-23-16216	3/9/2023	3/9/2025	TRUCK 15 - 1000 LB
GL1502	MI--03-23-16216	3/9/2023	3/9/2025	TRUCK 15 - 1000 LB
GL1503	MI--03-23-16216	3/9/2023	3/9/2025	TRUCK 15 - 1000 LB
GL1504	MI--03-23-16216	3/9/2023	3/9/2025	TRUCK 15 - 1000 LB
GL1505	MI--03-23-16216	3/9/2023	3/9/2025	TRUCK 15 - 1000 LB

Technician(s): LUCAS SULLIVAN;MATT PLEVA;

Signature: Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.