



# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0034432-1  
**Request Date** 4/12/2023 7:56:00 AM  
**Schedule Date** 4/14/2023 1:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1142  
 SUP1142  
 SUPERIOR MATERIALS 4903 (20)  
 G 5300 NORTH DORT HWY  
 FLINT , MI, 48505  
 (248) 788-8000

**Customer P.O** 6300426010  
**Requester** DAN THOMAS  
**Phone** (248) 788-8000  
**Email** dan.hargrave-  
 thomas@vcimentos.com

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK 1pm	
<b>WORK PERFORMED</b>	
CALL COMPLETE CALIBRATIONS COMPLETE	
Inspected, tested, calibrated and stickered cement and agg scale. Calibration reports to be done at the shop.	

Service Date 4/14/2023

**2 hours minimum portal to portal**

TECHNICIAN(S)
LUCAS SULLIVAN, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40000 lb x 20 lb	62822A	62822A	ALKON SCALE/INDICATOR / COMMAND	Aggregate
10000 lb x 5 lb	62822C	62822C	ALKON SCALE/INDICATOR / COMMAND	Cement

PARTS LIST

Customer Signature

\_\_\_\_\_  
Name:



# WORK ORDER

35380 UNION LAKE RD  
HARRISON TOWNSHIP, MI 48045  
Phone: (586)776-0517  
Fax: (586)776-3169  
Fed Tax ID: 38-3206225

**Work Order#** 0034432-1  
**Request Date** 4/12/2023 7:56:00 AM  
**Schedule Date** 4/14/2023 1:00:00 PM

**Bill To** SUP0896  
SUP0896  
SUPERIOR MATERIALS HOLDINGS LLC  
PO BOX 2900  
FARMINGTON HILLS, MI, 48336  
(248) 788-8000

**Ship To** SUP1142  
SUP1142  
SUPERIOR MATERIALS 4903 (20)  
G 5300 NORTH DORT HWY  
FLINT , MI, 48505  
(248) 788-8000

**Customer P.O** 6300426010  
**Requester** DAN THOMAS  
**Phone** (248) 788-8000  
**Email** dan.hargrave-  
thomas@vcimentos.com

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK 1pm	

Service Date \_\_\_\_\_

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE	LUCAS SULLIVAN, ROB DUBAY
Inspected, tested, calibrated and stickered cement and agg scale. Calibration reports to be done at the shop.	

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40000 lb x 20 lb	62822A	62822A	ALKON SCALE/INDICATOR / COMMAND	Aggregate
10000 lb x 5 lb	62822C	62822C	ALKON SCALE/INDICATOR / COMMAND	Cement

PARTS LIST

Customer Signature

Jeff



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 4903 (20)  
 SUP1142 G 5300 NORTH DORT HWY

**Certificate ID:** 62822C230421024547

**Test Date:** 04/14/2023

FLINT, MI 48505

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62822C Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/14/2023	10/14/2023	Semi-Annually	62822C	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Cement		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT
				5

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	1000	lb	1000	lb	0	lb
6000	lb	6000	lb	0	lb	2000	lb	2000	lb	0	lb
0	lb	0	lb	0	lb	3000	lb	3000	lb	0	lb
	lb		lb		lb	4000	lb	4000	lb	0	lb
	lb		lb		lb	5000	lb	5000	lb	0	lb
	lb		lb		lb	6000	lb	6000	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	

**OVERALL TEST RESULTS PASS**



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 4903 (20)  
 SUP1142 G 5300 NORTH DORT HWY

**Certificate ID:** 62822A230421024907

**Test Date:** 04/14/2023

FLINT , MI 48505

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62822A Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/14/2023	10/14/2023	Semi-Annually	62822A	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
Aggregate		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM lb	INCREMENT 20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS						
Weight Applied	Reading	Error	Weight Applied	Reading	Error						
1000	lb	1000	lb	0	lb	1000	lb	1000	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
3000	lb	3000	lb	0	lb	3000	lb	3000	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error				
3000	lb	3000	lb	6000	lb	0	lb
3000	lb	6000	lb	9000	lb	0	lb
3000	lb	9000	lb	12000	lb	0	lb
3000	lb	12000	lb	15000	lb	0	lb
3000	lb	15000	lb	18000	lb	0	lb
3000	lb	18000	lb	21000	lb	0	lb
3000	lb	21000	lb	24000	lb	0	lb
3000	lb	24000	lb	27000	lb	0	lb
3000	lb	27000	lb	30000	lb	0	lb
3000	lb	30000	lb	33000	lb	0	lb
3000	lb	33000	lb	36000	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 4903 (20)  
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

**Certificate ID:** 62822A230421024907

**Test Date:** 04/14/2023

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**  
 scale on zero, load tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** ROBERT DUBAY  Digitally signed by ROBERT DUBAY

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

**Customer:** SUPERIOR MATERIALS 4903 (20)  
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

**Certificate ID:** 62822C230421024547

**Test Date:** 04/14/2023

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

N/A

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL0601	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0602	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0603	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0604	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0606	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0605	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** ROBERT DUBAY Digitally signed by ROBERT DUBAY

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.