



# Metro Scale Company, Inc

17670 Allen Rd, Melvindale  
MI, 48122  
Phone: 313.382.6190  
Fax: 313.382.6194

## Certificate of Calibration

**Customer:** Superior Materials Plant # 21  
**Address:** 17200 Dix Toledo Rd  
**City, State Zip:** Brownstown, MICHIGAN 48193

**Certificate ID:** 23635-2262502  
**ISO Number:**  
**Date:** 2/25/2026

**Indicator Mfr.**  
COMMAND ALKON  
**Indicator Model**  
NA  
**Indicator Serial**  
23635-2  
**Test Interval**  
6 Months

**Base Mfr.**  
N/A  
**Base Model**  
N/A  
**Base Serial**  
N/A

**Cal Date**  
2/25/2026  
**Due Date**  
8/31/2026  
**Procedure**  
HB-44

**Scale ID**  
23635-2  
**Scale Class**  
III  
**Scale Status**  
In Service

**Scale Location**  
Agg Scale  
**Scale Range**  
0 - 50000 lb x 20 lb

### EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

### LOAD TEST

Preliminary Load Test			
Test Wt.	Tolerance	Reading	Error
0 lb	-20 - 20 lb	0 lb	0 lb
1000 lb	980 - 1020 lb	1000 lb	0 lb
2000 lb	1980 - 2020 lb	2000 lb	0 lb
3000 lb	2980 - 3020 lb	3000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

Final Load Test			
Test Wt.	Tolerance	Reading	Error
0 lb	-20 - 20 lb	0 lb	0 lb
1000 lb	980 - 1020 lb	1000 lb	0 lb
2000 lb	1980 - 2020 lb	2000 lb	0 lb
3000 lb	2980 - 3020 lb	3000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

### BUILDUP TEST

Build-up Test Result: [Pass] Fail Adjust Not Applicable

Location	Original Weight	Applied Weight	Total Weight	Reading	Error
	3240 lb	1000 lb	4240 lb	4240 lb	0 lb
	3240 lb	2000 lb	5240 lb	5240 lb	0 lb
	3240 lb	3000 lb	6240 lb	6240 lb	0 lb
	6000 lb	3000 lb	9000 lb	9000 lb	0 lb
	9020 lb	3000 lb	12020 lb	12020 lb	0 lb
	15420 lb	3000 lb	18420 lb	18420 lb	0 lb
	21200 lb	3000 lb	24200 lb	24200 lb	0 lb
	27120 lb	3000 lb	30120 lb	30120 lb	0 lb
	33400 lb	3000 lb	36400 lb	36400 lb	0 lb

### TEST INFORMATION

**Test Weight Classification:**

F

**Traceability Certificate Number(s):**

MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026

**Standards Used:**

101, 104, 105

**Expanded Uncertainty:**

9100 grams

**Test Location:**

[Onsite] Offsite

**Overall Result:**

[Pass] Fail Adjust

**Was the scale within customers required accuracy?**

[Y] N N/A

**Environmental Conditions:**

[Acceptable] Unacceptable

**Temperature:**

34°F

**Humidity:**

53%

**Comments / Notes:**

N/A

**Technician:**

Alex Lulko

**Certifying Technician:**

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.



# Metro Scale Company, Inc

17670 Allen Rd, Melvindale  
MI, 48122  
Phone: 313.382.6190  
Fax: 313.382.6194

## Certificate of Calibration

**Customer:** Superior Materials Plant # 21  
**Address:** 17200 Dix Toledo Rd  
**City, State Zip:** Brownstown, MICHIGAN 48193

**Certificate ID:** 23635-1262502  
**ISO Number:**  
**Date:** 2/25/2026

**Indicator Mfr.**  
COMMAND ALKON  
**Indicator Model**  
NA  
**Indicator Serial**  
23635-1  
**Test Interval**  
6 Months

**Base Mfr.**  
N/A  
**Base Model**  
N/A  
**Base Serial**  
N/A

**Cal Date**  
2/25/2026  
**Due Date**  
8/31/2026  
**Procedure**  
HB-44

**Scale ID**  
23635-1  
**Scale Class**  
III  
**Scale Status**  
In Service

**Scale Location**  
Cement Scale  
**Scale Range**  
0 - 10000 lb x 5 lb

### EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

### LOAD TEST

Preliminary Load Test		
Test Wt.	Reading	Error
0 lb	0 lb	0 lb
2000 lb	1995 lb	-5 lb
4000 lb	4000 lb	0 lb
6000 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

Final Load Test		
Test Wt.	Reading	Error
0 lb	0 lb	0 lb
2000 lb	1995 lb	-5 lb
4000 lb	4000 lb	0 lb
6000 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

### TEST INFORMATION

**Test Weight Classification:**  
**Traceability Certificate Number(s):**  
**Standards Used:**  
**Expanded Uncertainty:**  
**Test Location:**  
**Overall Result:**  
**Was the scale within customers required accuracy?**  
**Environmental Conditions:**  
**Temperature:**  
**Humidity:**  
**Comments / Notes:**  
**Technician:**

F  
MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026  
101, 102, 103, 104, 105, 106  
Available on Request or Reported on this Document  
[Onsite] Offsite  
[Pass] Fail Adjust  
[Y] N N/A  
[Acceptable] Unacceptable  
34°F  
53%  
N/A  
Alex Lulko

**Certifying Technician:**

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.