



# Metro Scale Company, Inc

17670 Allen Rd, Melvindale  
MI, 48122  
Phone: 313.382.6190  
Fax: 313.382.6194

## Certificate of Calibration

**Customer:** Superior Materials Plant # 36  
**Address:** 2470 Auburn Rd  
**City, State Zip:** Auburn Hills , MICHIGAN 48326

**Certificate ID:** CBS62803C250511  
**ISO Number:**  
**Date:** 11/5/2025

<b>Indicator Mfr.</b> COMMAND ALKON	<b>Base Mfr.</b> N/A	<b>Cal Date</b> 11/5/2025	<b>Scale ID</b> CBS62803C	<b>Scale Location</b> Cement Scale
<b>Indicator Model</b> NA	<b>Base Model</b> N/A	<b>Due Date</b> 5/31/2026	<b>Scale Class</b> III	<b>Scale Range</b> 0 - 10000 lb x 5 lb
<b>Indicator Serial</b> CBS62803C	<b>Base Serial</b> N/A	<b>Procedure</b> HB-44	<b>Scale Status</b> In Service	
<b>Test Interval</b> 2 Years				

### EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

### LOAD TEST

Preliminary Load Test		
Test Wt.	Reading	Error
0 lb	0 lb	0 lb
2000 lb	2000 lb	0 lb
4000 lb	4000 lb	0 lb
6000 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

Final Load Test		
Test Wt.	Reading	Error
0 lb	0 lb	0 lb
2000 lb	2000 lb	0 lb
4000 lb	4000 lb	0 lb
6000 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

### TEST INFORMATION

**Test Weight Classification:** F  
**Traceability Certificate Number(s):** MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026  
**Standards Used:** 101, 102, 103, 104, 105, 106  
**Expanded Uncertainty:** 2300 grams  
**Test Location:** [Onsite] Offsite  
**Overall Result:** [Pass] Fail Adjust  
**Was the scale within customers required accuracy?** [Y] N N/A  
**Environmental Conditions:** [Acceptable] Unacceptable  
**Comments / Notes:** N/A  
**Technician:** Alex Lulko

**Certifying Technician:**

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.



# Metro Scale Company, Inc

17670 Allen Rd, Melvindale  
MI, 48122  
Phone: 313.382.6190  
Fax: 313.382.6194

## Certificate of Calibration

**Customer:** Superior Materials Plant # 36  
**Address:** 2470 Auburn Rd  
**City, State Zip:** Auburn Hills , MICHIGAN 48326

**Certificate ID:** CBS62803A250511  
**ISO Number:**  
**Date:** 11/5/2025

**Indicator Mfr.**  
COMMAND ALKON  
**Indicator Model**  
NA  
**Indicator Serial**  
CBS62803A  
**Test Interval**  
2 Years

**Base Mfr.**  
N/A  
**Base Model**  
N/A  
**Base Serial**  
N/A

**Cal Date**  
11/5/2025  
**Due Date**  
5/31/2026  
**Procedure**  
HB-44

**Scale ID**  
CBS62803A  
**Scale Class**  
III  
**Scale Status**  
In Service

**Scale Location**  
Agg Scale A  
**Scale Range**  
0 - 40000 lb x 20 lb

### EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

### LOAD TEST

Preliminary Load Test			
Test Wt.	Tolerance	Reading	Error
0 lb	-20 - 20 lb	0 lb	0 lb
2000 lb	1980 - 2020 lb	2000 lb	0 lb
4000 lb	3980 - 4020 lb	4000 lb	0 lb
6000 lb	5980 - 6020 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

Final Load Test			
Test Wt.	Tolerance	Reading	Error
0 lb	-20 - 20 lb	0 lb	0 lb
2000 lb	1980 - 2020 lb	2000 lb	0 lb
4000 lb	3980 - 4020 lb	4000 lb	0 lb
6000 lb	5980 - 6020 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

### BUILDUP TEST

Build-up Test Result: [Pass] Fail Adjust Not Applicable

Location	Original Weight	Applied Weight	Total Weight	Reading	Error
	6100 lb	3000 lb	9100 lb	9100 lb	0 lb
	6100 lb	6000 lb	12100 lb	12100 lb	0 lb
	12240 lb	3000 lb	15240 lb	15240 lb	0 lb
	12240 lb	6000 lb	18240 lb	18240 lb	0 lb
	18000 lb	6000 lb	24000 lb	24000 lb	0 lb
	24160 lb	6000 lb	30160 lb	30160 lb	0 lb
	30040 lb	3000 lb	33040 lb	33040 lb	0 lb

### TEST INFORMATION

**Test Weight Classification:**

F

**Traceability Certificate Number(s):**

MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026

**Standards Used:**

101, 102, 103, 104, 105, 106

**Expanded Uncertainty:**

9100 grams

**Test Location:**

[Onsite] Offsite

**Overall Result:**

[Pass] Fail Adjust

**Was the scale within customers required accuracy?**

[Y] N N/A

**Environmental Conditions:**

[Acceptable] Unacceptable

**Comments / Notes:**

N/A

**Technician:**

Alex Lulko

**Certifying Technician:**

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.



# Metro Scale Company, Inc

17670 Allen Rd, Melvindale  
MI, 48122  
Phone: 313.382.6190  
Fax: 313.382.6194

## Certificate of Calibration

**Customer:** Superior Materials Plant # 36  
**Address:** 2470 Auburn Rd  
**City, State Zip:** Auburn Hills , MICHIGAN 48326

**Certificate ID:** CBS62803B250511  
**ISO Number:**  
**Date:** 11/5/2025

**Indicator Mfr.**  
COMMAND ALKON  
**Indicator Model**  
NA  
**Indicator Serial**  
CBS62803B  
**Test Interval**  
2 Years

**Base Mfr.**  
N/A  
**Base Model**  
N/A  
**Base Serial**  
N/A

**Cal Date**  
11/5/2025  
**Due Date**  
5/31/2026  
**Procedure**  
HB-44

**Scale ID**  
CBS62803B  
**Scale Class**  
III  
**Scale Status**  
In Service

**Scale Location**  
Agg Scale B  
**Scale Range**  
0 - 40000 lb x 20 lb

### EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

### LOAD TEST

Preliminary Load Test			
Test Wt.	Tolerance	Reading	Error
0 lb	-20 - 20 lb	0 lb	0 lb
2000 lb	1980 - 2020 lb	2000 lb	0 lb
4000 lb	3980 - 4020 lb	4000 lb	0 lb
6000 lb	5980 - 6020 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

Final Load Test			
Test Wt.	Tolerance	Reading	Error
0 lb	-20 - 20 lb	0 lb	0 lb
2000 lb	1980 - 2020 lb	2000 lb	0 lb
4000 lb	3980 - 4020 lb	4000 lb	0 lb
6000 lb	5980 - 6020 lb	6000 lb	0 lb

In maintenance tolerance? [Yes] No N/A

### BUILDUP TEST

Build-up Test Result: [Pass] Fail Adjust Not Applicable

Location	Original Weight	Applied Weight	Total Weight	Reading	Error
	6280 lb	3000 lb	9280 lb	9280 lb	0 lb
	6280 lb	6000 lb	12280 lb	12280 lb	0 lb
	11980 lb	3000 lb	14980 lb	14980 lb	0 lb
	11980 lb	6000 lb	17980 lb	17980 lb	0 lb
	18020 lb	6000 lb	24020 lb	24020 lb	0 lb
	23960 lb	6000 lb	29960 lb	29960 lb	0 lb
	30100 lb	3000 lb	33100 lb	33100 lb	0 lb

### TEST INFORMATION

**Test Weight Classification:**

F

**Traceability Certificate Number(s):**

MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026

**Standards Used:**

101, 102, 103, 104, 105, 106

**Expanded Uncertainty:**

9100 grams

**Test Location:**

[Onsite] Offsite

**Overall Result:**

[Pass] Fail Adjust

**Was the scale within customers required accuracy?**

[Y] N N/A

**Environmental Conditions:**

[Acceptable] Unacceptable

**Comments / Notes:**

N/A

**Technician:**

Alex Lulko

**Certifying Technician:**

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.