



Metro Scale Company, Inc

17670 Allen Rd, Melvindale
MI, 48122
Phone: 313.382.6190
Fax: 313.382.6194

Certificate of Calibration

Customer: Superior Materials Plant # 38
Address: 901 S. Wagner Rd
City, State Zip: Ann Arbor, MICHIGAN 48103

Certificate ID: 102559-1261501
ISO Number:
Date: 1/14/2026

| | | | | |
|--|---------------------------|------------------------------|-----------------------------------|---|
| Indicator Mfr. COMMAND ALKON | Base Mfr. N/A | Cal Date 1/14/2026 | Scale ID 102559-1 | Scale Location Cement Scale |
| Indicator Model NA | Base Model N/A | Due Date 7/31/2026 | Scale Class III | Scale Range 0 - 10000 lb x 5 lb |
| Indicator Serial 102559-1 | Base Serial N/A | Procedure HB-44 | Scale Status In Service | |
| Test Interval 2 Years | | | | |

EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

LOAD TEST

| Preliminary Load Test | | |
|-----------------------|---------|-------|
| Test Wt. | Reading | Error |
| 0 lb | 0 lb | 0 lb |
| 2000 lb | 2000 lb | 0 lb |
| 4000 lb | 4000 lb | 0 lb |
| 6000 lb | 6000 lb | 0 lb |
| 2000 lb | 2000 lb | 0 lb |
| 0 lb | 0 lb | 0 lb |

In maintenance tolerance? [Yes] No N/A

| Final Load Test | | |
|-----------------|---------|-------|
| Test Wt. | Reading | Error |
| 0 lb | 0 lb | 0 lb |
| 2000 lb | 2000 lb | 0 lb |
| 4000 lb | 4000 lb | 0 lb |
| 6000 lb | 6000 lb | 0 lb |
| 2000 lb | 2000 lb | 0 lb |
| 0 lb | 0 lb | 0 lb |

In maintenance tolerance? [Yes] No N/A

TEST INFORMATION

Test Weight Classification: F
Traceability Certificate Number(s): MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026
Standards Used: 101, 102, 103, 104, 105, 106
Expanded Uncertainty: 2300 grams
Test Location: [Onsite] Offsite
Overall Result: [Pass] Fail Adjust
Was the scale within customers required accuracy? [Y] N N/A
Environmental Conditions: [Acceptable] Unacceptable
Comments / Notes: N/A
Technician: Alex Lulko

Certifying Technician:

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.



Metro Scale Company, Inc

17670 Allen Rd, Melvindale
MI, 48122
Phone: 313.382.6190
Fax: 313.382.6194

Certificate of Calibration

Customer: Superior Materials Plant # 38
Address: 901 S. Wagner Rd
City, State Zip: Ann Arbor, MICHIGAN 48103

Certificate ID: 102559-2261501
ISO Number:
Date: 1/14/2026

Indicator Mfr.
COMMAND ALKON
Indicator Model
NA
Indicator Serial
102559-2
Test Interval
2 Years

Base Mfr.
N/A
Base Model
N/A
Base Serial
N/A

Cal Date
1/14/2026
Due Date
1/31/2028
Procedure
HB-44

Scale ID
102559-2
Scale Class
III
Scale Status
In Service

Scale Location
Agg Scale
Scale Range
0 - 40000 lb x 20 lb

EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

LOAD TEST

| Preliminary Load Test | | | |
|-----------------------|----------------|---------|-------|
| Test Wt. | Tolerance | Reading | Error |
| 0 lb | -20 - 20 lb | 0 lb | 0 lb |
| 2000 lb | 1980 - 2020 lb | 2000 lb | 0 lb |
| 4000 lb | 3980 - 4020 lb | 4000 lb | 0 lb |
| 6000 lb | 5980 - 6020 lb | 6000 lb | 0 lb |

In maintenance tolerance? [Yes] No N/A

| Final Load Test | | | |
|-----------------|----------------|---------|-------|
| Test Wt. | Tolerance | Reading | Error |
| 0 lb | -20 - 20 lb | 0 lb | 0 lb |
| 2000 lb | 1980 - 2020 lb | 2000 lb | 0 lb |
| 4000 lb | 3980 - 4020 lb | 4000 lb | 0 lb |
| 6000 lb | 5980 - 6020 lb | 6000 lb | 0 lb |

In maintenance tolerance? [Yes] No N/A

BUILDUP TEST

Build-up Test Result: [Pass] Fail Adjust Not Applicable

| Location | Original Weight | Applied Weight | Total Weight | Reading | Error |
|----------|-----------------|----------------|--------------|----------|-------|
| | 6020 lb | 6000 lb | 12020 lb | 12020 lb | 0 lb |
| | 12100 lb | 6000 lb | 18100 lb | 18100 lb | 0 lb |
| | 18420 lb | 6000 lb | 24420 lb | 24420 lb | 0 lb |
| | 24340 lb | 6000 lb | 30340 lb | 30340 lb | 0 lb |
| | 30000 lb | 6000 lb | 36000 lb | 36000 lb | 0 lb |

TEST INFORMATION

Test Weight Classification: F
Traceability Certificate Number(s): MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026
Standards Used: 101, 102, 103, 104, 105, 106
Expanded Uncertainty: 9100 grams
Test Location: [Onsite] Offsite
Overall Result: [Pass] Fail Adjust
Was the scale within customers required accuracy? [Y] N N/A
Environmental Conditions: [Acceptable] Unacceptable
Temperature: 24°F
Humidity: 81%
Comments / Notes: N/A
Technician: Alex Lulko

Certifying Technician:

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.